

Manufacturer's Certificate of Validation

Certificate No.	Date	Page
VC160200278	March 03,2016	1 of 1

imec Messtechnik GmbH certifies that the products mentioned on the following page(s) have been thoroughly tested, validated and met performance accuracy specifications over the stated ranges.

Reference Instrumentation

Device	Model	Calibration Date	Accuracy/Volatility	Serial Number
Constant Temperature				
Trough of Low Temperature	HC-2010	March 25,2015	± 0.05°C	JYDQ-135
Check Equipment				
Low-Temperature	ATL60	March 25,2015	± 0.015°C	JYDQ-136

All reference instruments are calibrated by SuZhou LonBo Calibration and Testing Co.,LTD, a CNAS (China National Accreditation Service) accredited laboratory, and traceable to NIST according to Multi-Recognition Agreement.

Environment Conditions

Air Temperature
25.5°C
Relative Humidity
53.0%

Validation Information

Validation Points	Tolerance	Result
+60°C	± 0.5°C	PASS
+40°C	± 0.5°C	PASS
+25°C	± 0.5°C	PASS
0°C	± 0.5°C	PASS
-5°C	± 0.5°C	PASS
-25°C	± 0.5°C	PASS

Product Information

Product Name	Serial Number	Production Date	Expiry Date
tempmate®-M1 V1.2	TMM160200278	03/2016	03/2017

This certificate was automatically generated and valid without signature for 1 (one) year from date of production. V10/2015.